503. Humidity

Item of Calibration	Field Code	Range	CMC (Confidence level approximately 95 %)	Comments
Relative humidity hygrometers;	50302			Referance Humidity/Thermometer Temperature/humidity chamber
Polimer Thinfilm (Humidity)		(20 ~ 60) %rh	2.4 %rh	/SRST-CI-50302-1
		(60 ~ 95) %rh	3.1 %rh	/SRST-CI-50302-2
Polimer Thinfilm (Temperature)		(-20 ~ 60) ℃	0.6 ℃	
Hair (Humidity)		(20 ~ 60) %rh	2.7 %rh	
		(60 ~ 95) %rh	3.3 %rh	
Hair (Temperature)		(-20 ~ 60) ℃	0.6 ℃	
Temperature humidity recorders ; Hygrothermograph, etc	50304			Referance Humidity/Thermometer Temperature/humidity chamber
Humidity		(20 ~ 60) %rh	2.7 %rh	/SRST-CI-50304
		(60 ~ 95) %rh	3.3 %rh	
Temperature		(-15 ~ 50) ℃	0.7 ℃	
Transducers; relative humidity	50305			Referance Humidity/Thermometer
Humidity		(20 ~ 60) %rh	2.2 %rh	Temperature/humidity chamber
		(60 ~ 95) %rh	3.0 %rh	/SRST-CI-50305
Humidity generators; constant temperature and humidity chamber	50306			Referance Humidity/Thermometer
Humidity		(20 ~ 40) %rh	2.2 %rh	/SRST-CI-50306
		(40 ~ 60) %rh	2.8 %rh	
		(60 ~ 80) %rh	3.3 %rh	
		(80 ~ 95) %rh	3.8 %rh	
Temperature		(-30 ~ 100) ℃	0.9 ℃	